

2607

ABSTRACT

Control apparatus for an ultrasonic phacoemulsification handpiece utilizes flow rates and temperature measurements as well as power provided to a handpiece for calculating a matrix of power level duty cycle combination that would not generate sufficient heat to create a burn in eye tissue and either prevent operation of handpiece in the matrix or alert a surgeon of such use.